

Measuring Salt Content in Chili Sauce with LAQUAtwin Salt-22 Pocket Meter

Chili sauces may contain high amount of salts that increase chances of kidney and heart diseases. To measure the salt content in chili sauce, simply dilute a weighed sample with distilled or deionized water and then place some drops onto the sensor of LAQUAtwin Salt-22 pocket meter. The meter measures the salt content accurately and displays the result in percentage (%) unit in just a few seconds.



Introduction

Chili sauce is a popular condiment that adds spice and flavour to food. It may be hot, sweet, or a combination thereof, and may have a thicker texture and viscosity compared to that of hot sauces. The ingredients of chili sauces vary, but typically include cooked chili peppers, vinegar, sugar, salt and sometimes red tomato. Some chili sauces available commercially are loaded with salt and other preservatives that can be damaging to your health.

Salt, chemically known as sodium chloride (NaCl), is a commonly used seasoning ingredient that serves as flavour enhancer and preservative in packaged or processed foods. It aids in balancing sweetness and suppressing bitterness. It also keeps the growth of pathogenic organisms at bay. Most bacteria, fungi, and other potentially pathogenic organisms cannot survive in a highly salty environment. Such environment is hypertonic, which will dehydrate any living cell causing it to die or temporarily inactivated.

The LAQUAtwin Salt-22 pocket meter offers fast, simple, and easy way of measuring salt content in packaged or processed foods such as chili sauces. This waterproof pocket meter with backlit LCD measures the sodium concentration of a microvolume

sample and converts it to salt or sodium chloride concentration in percentage (%) unit. The replaceable sensor has a sample well embedded with flat sodium ion selective sensor. This unique sensor can hold sample as little as 0.3ml or 0.05ml when used with sampling sheet.

Method

Meter Set-up and Calibration

Calibrate the meter according to manufacturer's instructions using 0.5% and 5.0% NaCl standards that come with the kit.

Sample Preparation and Measurement

1. Weigh a portion of the chili sauce and add distilled or deionized water.
2. Mix the diluted sample thoroughly.
3. Using a dropper, place some drops of diluted sample onto the sensor.
4. Record the stable reading.

Chili sauce should be diluted with distilled or deionized water to liberate the salt. The reading of the diluted sample should fall within the calibrated

measurement range of the meter. To obtain accurate results, a uniform temperature should be maintained for the standards and samples.

After measurement, clean the sensor with detergent and warm water. If there are still sample residues or stains after cleaning, place some drops of household bleach ($\leq 5\%$ sodium hypochlorite) onto the sensor and leave it for 5 to 30 minutes. Rinse the sensor with clean water and blot dry with soft tissue. For more information on maintenance, refer to Technical Tip 2: LAQUAtwin Ion Sensor Maintenance Procedures.

Results and Benefits

Chili sauces may contain varying amounts of salt. Salt is soluble in water. It contains 40% sodium (Na), which is a mineral required by the body in small amount for maintaining blood pressure and fluid balance as well as transmitting nerve impulses. Besides salt, sodium is also found in monosodium glutamate (MSG), baking soda, and baking powder. Too much sodium can increase blood pressure and damage kidneys.

For packaged or processed foods, sodium is listed as milligrams (mg) per serving as

Continued at the back

well as per 100g on the nutrition information panel of the label, rather than salt. Table 1 shows the measurement results of chili sauce with LAQUAtwin Salt-22 pocket meter.

Table 1: Measurement Results of Chili Sauce with LAQUAtwin Salt-22 Pocket Meter

A	Sodium (mg / 100g) on Bottle Label	4520
B	Chili Sauce Weight (g)	10.0618
C	Volume (ml) or weight (g) After Dilution	100
D	LAQUAtwin Salt-22 Pocket Meter Reading (% NaCl)	1.1
E	Calculated NaCl Salt (g / 100g)	10.93
F	Calculated Na (mg / 100g)	4300
G	Recovery (%)	95

To calculate the amount of sodium from salt (NaCl) reading obtained with LAQUAtwin Salt-22 pocket meter, use the equations in Table 2. Refer to the values in Table 1 to follow the sample calculations.

Table 2: Sodium Chloride (NaCl) and Sodium Calculations

Meter Reading	1.1% NaCl = 1.1g NaCl / 100 ml or g dil. chili sauce
NaCl Salt (g / 100g)	Meter reading (D) x Dilution Factor (C/B) x 100
	$\frac{1.1\text{g NaCl}}{100\text{ml (g)}} \times \frac{100\text{ml (g)}}{10.0618\text{g}} \times 100\text{g}$
Na (mg / 100g)	NaCl salt per 100g (E) x $\frac{\text{Na molar mass}}{\text{NaCl molar mass}} \times 1000$
	$10.93\text{g NaCl} \times \frac{22.99\text{ Na}}{58.44\text{ NaCl}} \times \frac{1000\text{mg}}{\text{g}}$
Recovery (%)	$\frac{\text{Calculated Sodium mg / 100g (F)}}{\text{Listed Sodium mg / 100g on Label (A)}} \times 100$
	$\frac{4300}{4520} \times 100$

According to the National Health Group Pharmacy in Singapore, adults should limit the amount of salt to one teaspoon per day, which is about 5g of salt or 2g of sodium. To ensure that you are not taking more than that amount, always check the amount of sodium per serving on the nutrition information panel of your chili sauce and other packaged or processed food. If that information is not available, simply perform a quick salt test with LAQUAtwin Salt-22 pocket meter.

References and Suggested Readings

1. Chili Sauce. Wikipedia. https://en.wikipedia.org/wiki/Chili_sauce
2. National Healthcare Group Pharmacy. All About Salt. https://www.pharmacy.nhg.com.sg/All_About_Salt/

REV 0, 19 July 2017

LAQUAtwin Pocket Ion Meters Lineup



HORIBA Instruments (Singapore) Pte. Ltd.
83 Science Park Drive, #02-02A, The Curie, Singapore 118258
Phone: 65 6908-9660
Fax: 65 6745-8155
www.horiba-laqua.com
e-mail: laqua@horiba.com

HORIBA UK Limited
Kyoto Close, Moulton Park, Northampton NN3 6FL
Phone: 44 (0) 1604 542567
Fax: 44 (0) 1604 542699
www.horiba.com/uk
e-mail: waterquality@horiba.com



www.horiba-laqua.com

IMS
HORIBA Group is operating Integrated Management System (IMS)
ISO9001 JOA-0298 / ISO14001
JOA-E-90039 / ISO13485
JOA-MD0010 / OHSAS18001 JOA-OH0068Rorum perid novis esimoente,

